

HIGH PRECISION BASE PLATE-TO-COUCH TOP FIXATION DEVICES 2-PIN BAR FOR PRODIGY™ COUCH (MRI COMPATIBLE)



Article No. : **32165**

A. GENERAL PRODUCT INFORMATION


This product is a fixation device to attach High Precision base plates to 530 mm wide couch tops with the Prodigy™ type indexing.

B. PRODUCT DESCRIPTION

This 2-pin bar has two positioning blocks that fit over the bulbs on the sides of Prodigy™ style couches. It can be used to index hardware (base plates, shoulder positioning device, leg indexing system, etc.) directly or via an adapter plate to the couch. For information on the compatible products, please refer to the respective 'instructions for use'.

C. POSITIONING ON A COUCH

The two blocks of this fixation device have a circular recess on the inside and fit perfectly over the bulbs of the Prodigy™ style indexing system.



Push the bar over a set of bulbs until it clicks in place. The bar itself should now be flat on the surface of the couch. Always make sure that the bar is well positioned on the couch before indexing the other hardware.

To remove the bar, hold the two ends of the bar and pull until the bar snaps off the couch. Never pull in the middle of the bar as this will push the positioning blocks even harder onto the bulbs of the couch.

D. STORAGE

Always store the product in a safe place to prevent it from getting damaged.

Store the product between +10°C (50°F) and 40°C (122°F).

E. PROPERTIES

E.1 Physical Properties

Dimensions : L 550 mm x W 25 mm x H 25 mm

Weight: 0.1 kg

E.2 Material Properties

Material POM (pins): 1,4 gr/cm³

Note:

It is prohibited to make alterations to this text without prior approval from Orfit Industries.

Material PA (blocks): 1,15 gr/cm³
Material PC (screws): 1,2 gr/cm³
Material Glass Fibre (bar): 1,8 gr/cm³
This product is made entirely of durable plastics and is MR-safe.



Warning: keep metal wrenches which might be supplied in the package out of the MRI room.

E.3 Dosimetric Properties

Attenuation factor (± 0.15 %)	
6 MV	2.67%
15 MV	2.00%
H ₂ O equivalence	7.72 mm

Note: Use these numbers as a guidance only. Perform the measurements again in your department to verify these results.

F. MAINTENANCE AND WASTE MANAGEMENT

This product can be cleaned and disinfected by means of soapy water or an isopropanol based disinfectant, applied with a soft cloth. If unsure about the cleaning fluid, do not use. **Never use aerosol sprays, corrosive cleaning agents, solvents or abrasive detergents.**

Periodic checks of the product should be done to insure the parts are not worn and require repair or replacement. **Do not attempt to make repairs yourself.** Contact your distributor if there are any questions or concerns.

The product can be disposed of with household waste.

G. FIXATION DEVICES COMBINATION PACKAGES

This product is part of following fixation devices combination packages:

- 32182
- 32185

H. ADDITIONAL INFORMATION

For additional information such as distributor contact information, product brochures, Safety Data Sheets and regulatory information, please visit our website www.orfit.com



ORFIT INDUSTRIES
Vosveld 9A | B-2110 Wijnegem | Belgium
T (+32) (0)3 326 20 26
welcome@orfit.com

www.orfit.com



Ref. No. 31989
VERSION 9
LAST UPDATE: 11/02/2025
REVISION DATE: 11/02/2027